

(19)



Europäisches Patentamt
European Patent Office
Office européen des brevets



(11)

EP 1 138 783 A2

(12)

EUROPEAN PATENT APPLICATION

(43) Date of publication:
04.10.2001 Bulletin 2001/40

(51) Int Cl.7: **C12Q 1/68, G06F 19/00**

(21) Application number: **01107737.7**

(22) Date of filing: **02.04.2001**

(84) Designated Contracting States:
**AT BE CH CY DE DK ES FI FR GB GR IE IT LI LU
MC NL PT SE TR**
Designated Extension States:
AL LT LV MK RO SI

(30) Priority: **31.03.2000 EP 00107036**
13.07.2000 DE 10034209

(71) Applicant: **Roche Diagnostics GmbH**
68298 Mannheim (DE)

(72) Inventors:
• **Tabliti, Karl, Dr.**
82343 Poecking (DE)
• **Sagner, Gregor, Dr.**
82377 Penzberg (DE)
• **Soong, Richie, Dr.**
82327 Tutzing (DE)
• **Gutekunst, Martin, Dr.**
82390 Eberfing (DE)

(54) **Method for the efficiency-corrected real-time quantification of nucleic acids**

(57) The present invention concerns a method for the quantification of a target nucleic acid in a sample comprising the following steps: (i) determination of the amplification efficiency of the target nucleic acid under defined amplification conditions, (ii) amplification of the target nucleic acid contained in the sample under the same defined reaction conditions, (iii) measuring the amplification in real-time, (iv) quantification of the origi-

nal amount of target nucleic acid in the sample by correction of the original amount derived from step (iii) with the aid of the determined amplification efficiency. The efficiency correction of PCR reactions according to the invention for the quantification of nucleic acids can be used for absolute quantification with the aid of an external or internal standard as well as for relative quantification compared to the expression of housekeeping genes.

EP 1 138 783 A2

Description

[0001] The present invention relates to the field of nucleic acid quantification with the aid of quantitative real-time PCR.

5 Prior art

[0002] Methods for the quantification of nucleic acids are important in many areas of molecular biology and in particular for molecular diagnostics. At the DNA level such methods are used for example to determine the copy numbers of gene sequences amplified in the genome. However, methods for the quantification of nucleic acids are used especially in connection with the determination of mRNA quantities since this is usually a measure for the expression of the respective coding gene.

[0003] If a sufficient amount of sample material is available, special mRNAs can be quantified by conventional methods such as Northern Blot analysis or RNase protection assay methods. However, these methods are not sensitive enough for sample material that is only available in small amounts or for genes that are expressed very weakly.

[0004] The so-called RT-PCR is a much more sensitive method. In this method a single-stranded cDNA is firstly produced from the mRNA to be analysed using a reverse transcriptase. Subsequently a double-stranded DNA amplification product is generated with the aid of PCR.

[0005] A distinction is made between two different variants of this method:

- In the so-called relative quantification the ratio of the expression of a certain target RNA is determined relative to the amount of RNA of a so-called housekeeping gene which is assumed to be constitutively expressed in all cells independent of the respective physiological status. Hence the mRNA is present in approximately the same amount in all cells.

The advantage of this is that different initial quantities of the various sample materials and the process of RNA preparation has no influence on the particular result. However, an absolute quantification is not possible with this method.

- Alternatively the absolute amount of RNA used can be determined with the aid of standard nucleic acids of a known copy number and amplification of a corresponding dilution series of this standard nucleic acid. There are two alternatives:

[0006] When using external standards the standard and target nucleic acid are amplified in separate reaction vessels. In this case a standard can be used with an identical sequence to the target nucleic acid. However, systematic errors can occur in this type of quantification if the RNA preparation to be analysed contains inhibitory components which impair the efficiency of the subsequent PCR reaction. Such errors can be excluded by using internal standards i.e. by amplifying the standard and target nucleic acid in one reaction vessel. However, a disadvantage of this method is that standards have to be used that have different sequences compared to the target nucleic acid to be analysed in order to be able to distinguish between the amplification of the standard and target nucleic acid. This can in turn lead to a systematic error in the quantification since different efficiencies of the PCR amplification cannot be excluded when the sequences are different.

[0007] PCR products can be quantified in two fundamentally different ways:

a) End point determination of the amount of PCR product formed in the plateau phase of the amplification reaction

In this case the amount of PCR product formed does not correlate with the amount of the initial copy number since the amplification of nucleic acids at the end of the reaction is no longer exponential and instead a saturation is reached. Consequently different initial copy numbers exhibit identical amounts of PCR product formed. Therefore the competitive PCR or competitive RT-PCR method is usually used in this procedure. In these methods the specific target sequence is coamplified together with a dilution series of an internal standard of a known copy number. The initial copy number of the target sequence is extrapolated from the mixture containing an identical PCR product quantity of standard and target sequence (Zimmermann and Mannhalter, Bio-Techniques 21:280-279, 1996). A disadvantage of this method is also that measurement occurs in the saturation region of the amplification reaction.

b) Kinetic real-time quantification in the exponential phase of PCR.

In this case the formation of PCR products is monitored in each cycle of the PCR. The amplification is usually measured in thermocyclers which have additional devices for measuring fluorescence signals during the amplification reaction. A typical example of this is the Roche Diagnostics LightCycler (Cat. No. 2 0110468). The amplification products are for example detected by means of fluorescent labelled hybridization probes which only emit fluorescence signals when they are bound to the target nucleic acid or in certain cases also by means of fluorescent

dyes that bind to double-stranded DNA. A defined signal threshold is determined for all reactions to be analysed and the number of cycles C_p required to reach this threshold value is determined for the target nucleic acid as well as for the reference nucleic acids such as the standard or housekeeping gene. The absolute or relative copy numbers of the target molecule can be determined on the basis of the C_p values obtained for the target nucleic acid and the reference nucleic acid (Gibson et al., Genome Research 6:995-1001; Bieche et al., Cancer Research 59:2759-2765, 1999; WO 97/46707; WO 97/46712; WO 97/46714).

[0008] In summary in all the described methods for the quantification of a nucleic acid by PCR the copy number formed during the amplification reaction is always related to the copy number formed of a reference nucleic acid which is either a standard or an RNA of a housekeeping gene. In this connection it is assumed that the PCR efficiency of the target and reference nucleic acid are not different.

[0009] Usually a PCR efficiency of 2.00 is assumed which corresponds to a doubling of the copy number per PCR cycle (User Bulletin No. 2 ABI Prism 7700, PE Applied Biosystems, 1997).

[0010] However, it has turned out that the real PCR efficiency can be different from 2.00 since it is influenced by various factors such as the binding of primers, length of the PCR product, G/C content and secondary structures of the nucleic acid to be amplified and inhibitors that may be present in the reaction mixture as a result of the sample preparation. This is particularly relevant when using heterologous reference nucleic acids e.g. in the relative quantification compared to the expression of housekeeping genes.

[0011] The object of the present invention was therefore to provide methods for the quantification of nucleic acids which overcome the disadvantages of the prior art as described above. The object of the present invention was in particular to provide methods for the quantification of nucleic acids in which a target nucleic acid is quantified independent of the amplification efficiencies of target nucleic acid and reference nucleic.

Brief description of the invention

[0012] This object is achieved according to the invention by a method for the quantification of a target nucleic acid in a sample comprising the following steps:

- a) Determining the amplification efficiency of the target nucleic acid under defined conditions.
- b) Amplifying the target nucleic acid contained in the sample under the same reaction conditions.
- c) Measuring the amplification in real-time.
- d) Quantifying the original amount of target nucleic acid in the sample by correction of the original amount derived from step c) with the aid of the determined amplification efficiency.

[0013] According to the invention this method can be used for relative quantification compared to the expression of housekeeping genes as well as for absolute quantification.

[0014] A first aspect of the invention therefore concerns a method for quantifying a target nucleic acid in a sample compared to a reference nucleic acid comprising the following steps:

- a) Determining the amplification efficiencies of the target nucleic acid and reference nucleic acid under defined amplification conditions
- b) Amplifying the target nucleic acid contained in the sample as well as the reference nucleic acid contained in the sample under the same defined amplification conditions.
- c) Measuring the amplification of the target nucleic acid and reference nucleic acid in real-time
- d) Calculating the original ratio of target nucleic acid and reference nucleic acid in the sample by correcting the ratio derived from step c) with the aid of the amplification efficiencies determined in step a).

[0015] A second aspect of the present invention concerns a method for the quantification of a target nucleic acid in a sample comprising the following steps:

- a) Determining the amplification efficiencies of the target nucleic acid and of an internal or an external standard under defined amplification conditions
- b) Amplifying the target nucleic acid contained in the sample as well as the internal or external standard under the same defined reaction conditions
- c) Measuring the amplification of the target nucleic acid and standard in real-time
- d) Calculating the original copy number in the sample by correcting the copy number derived from step c) with the aid of the amplification efficiencies determined in step a).

[0016] In all methods the amplification efficiencies are preferably determined by

- a) preparing a dilution series of the target nucleic acid
- b) amplifying the target nucleic acid under defined reaction conditions according to A, the amplification of the nucleic acids being measured in real-time
- c) setting a defined signal threshold value
- d) determining the cycle number for each dilution at which the signal threshold value is exceeded,
- e) calculating the amplification efficiency based on the determined cycle numbers.

Detailed description of the invention

[0017] The object of the invention is achieved by a method for the quantification of a target nucleic acid in a sample comprising the following steps:

- a) Determining the amplification efficiency of the target nucleic acid under defined conditions
- b) Amplifying the target nucleic acid contained in the sample under the same reaction conditions.
- c) Measuring the amplification in real-time
- d) Quantifying the original amount of target nucleic acid in the sample by correcting the original amount derived from step c) with the aid of the determined amplification efficiency.

[0018] The importance of an efficiency correction will be illustrated by an error calculation. Table 1 shows a theoretical calculation of the average percentage error of the determined copy number in the case of amplification efficiencies that are different from 2.00 as a function of the respective cycle number. The error is calculated according to the formula

$$\text{percentage error} = (2^n/E^n - 1) \times 100$$

in which E is the efficiency of the amplification and n is the respective cycle number at which the percentage error is determined.

Table 1:

Detection Cycle (n) PCR efficiency (E)	10	15	20	25	30	35
2.00	-	-	-	-	-	-
1.97	16%	25%	35%	46%	57%	70%
1.95	29%	46%	66%	88%	113%	142%
1.90	67%	116%	179%	260%	365%	500%
1.80	187%	385%	722%	1290%	2260%	3900%
1.70	408%	1045%	2480%	5710%	13.000%	29.500%
1.60	920%	2740%	8570%	26.400%	80.700%	246.400%

[0019] The amplification efficiency can be determined by various methods for example by determining a function with which the measured signal is determined relative to the amplification of the target nucleic acid as a function of the cycle time.

[0020] The amplification efficiency is preferably determined by a method in which

- a) a dilution series of a target nucleic acid is prepared
- b) the target nucleic acid is amplified under defined reaction conditions as claimed in claim 1 and the amplification of the nucleic acid is measured in real-time
- c) a defined signal threshold value is set
- d) for each dilution the cycle number C_p is determined at which the signal threshold value is exceeded
- e) a logarithmic linear function of the copy number of target nucleic acid used for the amplification is determined as a function of the cycle number at which the signal threshold value is exceeded
- f) the amplification efficiency E is calculated according to

$$E=G^{-a}$$

wherein a is determined as the first derivative of the function determined in step e) and G is the base number of the logarithm.

[0021] In a similar manner the amplification efficiency can also be determined by a method in which

- a) a dilution series of the target nucleic acid is prepared
- b) the target nucleic acid is amplified under defined reaction conditions as claimed in claim 1 and the amplification of the nucleic acid is measured in real-time
- c) a defined signal threshold value is set
- d) the cycle number Cp at which the signal threshold value is exceeded is determined for each dilution
- e) a linear function of the cycle number determined in step d) is determined as a function of a logarithm of the copy number of target nucleic acid used for the amplification and
- f) the amplification efficiency E is calculated according to

$$E=G^{-1/a}$$

wherein a is determined as the first derivative of the function determined in step e) and G is the base number of the logarithm.

[0022] Both preferred procedures have the advantage that a systematic error cannot occur that results from determining the amplification efficiency in a phase of the PCR reaction in which there is no longer an exponential amplification of the target nucleic acid (plateau phase).

[0023] However, it unexpectedly turned out that under certain conditions the amplification efficiency can also be dependent on the original amount of target nucleic acid or it can change during the first cycles of an amplification reaction that is still in the exponential phase. A subject matter of the invention is thus also a method for the efficiency-corrected quantification of nucleic acids in which the efficiency of the amplification is determined by

- a) preparing a dilution series of the target nucleic acid
- b) amplifying the target nucleic acid under defined reaction conditions as claimed in claim 1 and measuring the amplification of the nucleic acid in real-time
- c) setting a defined signal threshold value
- d) determining the cycle number Cp at which the signal threshold value is exceeded for each dilution
- e) determining the amplification efficiency as a function of the amount of target nucleic acid.

[0024] This can for example be achieved by derivation of a continuously differentiable function F(Cp) of the Cp values as a function of the original copy number or vice versa.

[0025] The function $F(Cp)=\log(\text{concentration of the original copy number})$ can for example be standardized by mathematical algorithms such as a polynomial fit of a higher degree. The amplification efficiency E can then be determined by the equation

$$E = G^{-dF(Cp)/dCp}$$

in which $dF/(Cp)$ is the derivative of the continuous function and G is the base number of the logarithm. A polynomial fit of the 4th degree has proven to be particularly suitable within the sense of the invention.

[0026] The efficiency-corrected quantification of nucleic acids according to the invention can in principle be used for methods for absolute quantification as well as for methods for relative quantification.

[0027] Hence a subject matter of the present invention in relation to relative quantification is also a method for the quantification of a target nucleic acid in a sample relative to a reference nucleic acid comprising the following steps:

- a) Determination of the amplification efficiencies of the target nucleic acid and of the reference nucleic acid under defined amplification conditions.
- b) Amplification of the target nucleic acid contained in the sample as well as of the reference nucleic acid contained in the sample under the same defined amplification conditions.

- c) Measurement of the amplification of the target nucleic acid and of the reference nucleic acid in real-time.
d) Calculation of the original ratio of target nucleic acid and reference nucleic acid in the sample by correcting the ratio derived from step c) with the aid of the amplification efficiencies determined in step a).

[0028] Such a method according to the invention eliminates on the one hand the influence of inhibitors that may be present in the examined sample and, on the other hand, corrects errors which may occur as a result of different amplification efficiencies of the target nucleic acid and reference nucleic acid.

[0029] Steps b) to d) are advantageously carried out in a parallel mixture containing a so-called calibrator sample. The calibrator sample is a sample which contains the target nucleic acid and reference nucleic acid in a defined ratio that is constant for each measurement. Subsequently the ratio of the quotients determined for the sample and for the calibrator sample is determined as a measure for the original amount of target nucleic acid in the sample. This has the advantage that in addition other systematic errors are eliminated that are due to differences in the detection sensitivity of the target nucleic acid and reference nucleic acid. Such systematic errors can for example occur as a result of different hybridization properties of the hybridization probes or, in the case of fluorescent-labelled probes, different excitation efficiencies, quantum yields or coupling efficiencies of the dye to the probe. Therefore the sample to be tested and the calibrator sample must be analysed in each experiment with the same detection agents i.e. with the same batch of fluorescent-labelled hybridization probes.

[0030] A special embodiment of relative quantification according to the invention is a method for the quantification of a target nucleic acid in a sample relative to a reference nucleic acid comprising the following steps:

- a) Determination of the amplification efficiencies of the target nucleic acid and of the reference nucleic acid under defined amplification conditions
- b) Amplification of the target nucleic acid contained in the sample and of the reference nucleic acid contained in the sample under the same defined amplification conditions.
- c) Measurement of the amplification of the target nucleic acid and of the reference nucleic acid in real time.
- d) Determination of a defined signal threshold value.
- e) Determination of the cycle numbers at which the signal threshold value is in each case exceeded during the amplification of the target nucleic acid and the reference nucleic acid.
- f) Calculation of the original ratio of target nucleic acid and reference nucleic acid in the sample according to the formula

$$N(T)_0/N(R)_0 = E(R)^{n(R)} / E(T)^{n(T)},$$

wherein

- $N(T)_0$ = the original amount of target DNA present in the sample
 $N(R)_0$ = the original amount of reference DNA present in the sample
 $E(R)$ = the amplification efficiency of the reference nucleic acid
 $n(R)$ = the cycle number of the reference nucleic acid measured in step e)
 $E(T)$ = the amplification efficiency of the target nucleic acid
 $n(T)$ = the cycle number of the target nucleic acid measured in step e)

[0031] In this embodiment it is advantageous to carry out steps b), c), e) and f) with a calibrator sample in order to eliminate systematic errors due to the detection of amplification products and subsequently the ratio of the quotients measured for the sample and for the calibrator sample are determined as a measure for the original amount of target nucleic acid in the sample.

[0032] The ratio obtained in step f) is calculated according to the invention as follows:

$$N(T)_n = N(T)_0 \times E(T)^{n(T)} \quad (1)$$

$$N(R)_n = N(R)_0 \times E(R)^{n(R)} \quad (2)$$

in which $N(T)_n$ = the amount of target DNA at the signal threshold value
and $N(R)_n$ = the amount of reference DNA at the signal threshold value

[0033] From (1) and (2) it follows that:

$$\frac{N(T)_n}{N(R)_n} = \frac{N(T)_0 \times E(T)^{n(T)}}{N(R)_0 \times E(R)^{n(R)}} \quad (3)$$

[0034] From this it follows that:

$$\frac{N(T)_0}{N(R)_0} = \frac{N(T)_n \times E(R)^{n(R)}}{N(R)_n \times E(T)^{n(T)}} \quad (4)$$

[0035] Due to the fact that an identical signal threshold value has been set for the target and reference nucleic acid this may be approximated to:

$$N(T)_n = N(R)_n.$$

[0036] Under this condition and starting from equation (4) for the original ratio of target nucleic acid and reference nucleic acid, this results in the equation

$$N(T)_0 / N(R)_0 = E(R)^{n(R)} / E(T)^{n(T)} \quad (5)$$

[0037] However, this assumed approximation does not apply when target nucleic acid and reference nucleic acid are detected with different sensitivities. According to the invention it is therefore particularly advantageous to measure a calibrator sample in a parallel reaction and to determine the ratio of the quotients $N(T)_0/N(R)_0$ measured for the sample and for the calibrator sample as a measure for the original amount of target nucleic acid in the sample.

[0038] This results in the following from equation (4) using the indices

_A for the sample to be analysed and
_K for the calibrator sample

$$\frac{\frac{N(T)_{0A}}{N(R)_{0A}}}{\frac{N(T)_{0K}}{N(R)_{0K}}} = \frac{\frac{N(T)_{nA} \times E(R)^{nA(K)}}{N(R)_{nA} \times E(T)^{nA(I)}}}{\frac{N(T)_{nK} \times E(R)^{nK(K)}}{N(R)_{nK} \times E(T)^{nK(I)}}} \quad (6)$$

[0039] Due to the fact that an identical signal threshold value has been set for the sample to be analysed and for the calibrator sample and that identical agents are used to detect target and reference amplicons in the sample and in the calibrator sample, the ratio of the quotient determined for the sample and for the calibrator sample is as follows:

$$\frac{N(T)_{nA}}{N(R)_{nA}} \bigg/ \frac{N(T)_{nK}}{N(R)_{nK}} = 1$$

[0040] Hence the ratio of the quotients of the sample to be analysed and the calibrator sample is:

$$\frac{N(T)_{0A}}{N(R)_{0A}} \bigg/ \frac{N(T)_{0K}}{N(R)_{0K}} = E(R)^{nA(K)-nK(K)*} E(T)^{nK(I)-nA(I)} \quad (7)$$

[0041] Consequently a relative value can be obtained in this manner for the original copy number of target nucleic acid in the sample in which systematic errors due to different amplification efficiencies as well as due to different detection sensitivities have been eliminated. The only requirement for the accuracy of the determined value is the justified assumption that under absolutely identical buffer conditions the amplification and detection efficiencies are also identical in the various reaction vessels.

[0042] Requirement for all methods according to the invention for relative quantification is that the amplification efficiency of the target nucleic acid as well as the amplification efficiency of the reference nucleic acid are determined. Both of these determinations are preferably carried out by the methods described above by determining the cycle number at which a certain signal threshold value is exceeded.

[0043] In a preferred embodiment of relative quantification the sample is divided into two aliquots and the real-time measurement of the amplification of the target nucleic acid and reference nucleic acid is carried out in separate reaction vessels. This prevents interference between the amplification reactions of the target nucleic acid and the reference nucleic acid with regard to their efficiency for example by competition for deoxynucleotides or Taq polymerase. Furthermore the target nucleic acid and reference nucleic acid can be detected with the same detection systems, for example with the same DNA binding dye.

[0044] Alternatively the real-time measurement of the amplification of target nucleic acid and reference nucleic acid can be carried out in one sample in the same reaction vessel using differently labelled hybridization probes. This is particularly advantageous when only small amounts of sample material are available because the number of PCR reactions required is halved in this manner.

[0045] If it is intended to determine the absolute amount of target nucleic acid to be detected in a sample, then the method for the quantification of a target nucleic acid in a sample comprises the following steps

- a) Determination of the amplification efficiencies of the target nucleic acid and of an internal or external standard under defined amplification conditions.
- b) Amplification of the target nucleic acid contained in the sample as well as of the internal or external standard under the same defined reaction conditions.
- c) Measurement of the amplification of the target nucleic acid and standard in real time
- d) Calculation of the original copy number in the sample by correcting the copy number derived from step c) with the aid of the amplification efficiencies determined in step a).

[0046] The sequences of the target nucleic acid and standard nucleic acid are advantageously substantially identical. However, when selecting the sequence for an internal standard it must be taken into account that the available detection system should be able to distinguish between the standard and target nucleic acid. This can for example be achieved by using hybridization probes with different labels for the detection of the target nucleic acid and internal standard. Ideally oligonucleotides are used for this as detection probes which can be used to distinguish between minimal sequence differences such as point mutations.

[0047] An advantage of using an internal standard is that the inhibitors present in the sample also influence the amplification of the standard. Hence differences in the amplification efficiencies can be minimized.

[0048] In contrast the use of an external standard has the advantage that the amplification reactions of the target nucleic acid and standard cannot competitively interfere with one another with regard to their efficiency. Moreover the amplification products of the standard and target nucleic acid can be detected in parallel reactions with the aid of the same detection system for example with the same hybridization probe. A disadvantage is possible differences in the PCR efficiencies due to inhibitors in the sample. However, errors in the quantification caused by this can be eliminated by the method described in the following:

[0049] In a preferred embodiment for the absolute quantification of a target nucleic acid in a sample the method according to the invention comprises the following steps:

- a) Determination of the amplification efficiencies of the target nucleic acid as well as of an internal or external standard under defined amplification conditions
- b) Amplification of the target nucleic acid contained in the sample as well as of the internal or external standard

under the same defined reaction conditions

c) Measurement of the amplification of target nucleic acid and standard in real-time

d) Setting a defined signal threshold value

e) Determination of the cycle number during the amplification of target nucleic acid and standard at which the signal threshold value is exceeded

f) Determination of the original copy number $N(T)_0$ of the target nucleic acid in the sample according to the formula

$$N(T)_0 = N(S)_0 \cdot E(S)^{n(S)} / E(T)^{n(T)} \quad (8)$$

in which

$N(S)_0$ = the original amount of standard used

$E(S)$ = the amplification efficiency of the standard

$n(S)$ = the cycle number of the standard measured in step e)

$E(T)$ = the amplification efficiency of the standard

$n(T)$ = the cycle number of the target nucleic acid measured in step e).

[0050] In this case like the relative quantification, the amplification efficiencies of the target nucleic acid and the internal standard are preferably determined as described by determining the cycle number at which a certain signal threshold value is exceeded.

[0051] According to the invention $N(T)_0$ is calculated as follows:

$$N(T)_n = N(T)_0 \cdot E(T)^{n(T)}$$

and

$$N(S)_n = N(S)_0 \cdot E(S)^{n(S)}$$

[0052] Since an identical signal threshold value has been set for the target and standard nucleic acid this approximates to:

$$N(T)_n = N(S)_n$$

[0053] Hence the original copy number of target nucleic acid present in the sample is calculated according to the equation

$$N(T)_0 = N(S)_0 \cdot E(S)^{n(S)} / E(T)^{n(T)} \quad (8)$$

[0054] The invention in particular also concerns those embodiments of the described methods for the efficiency-corrected quantification of nucleic acids in which the amplification products are detected by hybridization probes which can be labelled with a detectable component in many different ways.

[0055] A prerequisite for the efficiency-corrected determination of the original amount of a target nucleic acid and for the determination of the amplification efficiencies per se is to define signal threshold values and subsequently determine the cycle number for the respective amplification reaction at which a certain signal threshold value is reached.

The signal threshold value can be determined according to the prior art in various ways:

[0056] According to the prior art the signal threshold value can for example be a signal which corresponds to a certain multiple of the statistical variance of the background signal (ABI Prism 7700 Application Manual, Perkin Elmer).

[0057] Alternatively the cycle number at which the signal threshold value is exceeded can be determined according to the so-called "fit point above threshold" method (LightCycler Operator's Manual, B59-B68, Roche Molecular Biochemicals, 1999).

[0058] In a further embodiment the threshold value can be determined as a relative value instead of an absolute value when, independently of the absolute value of the signal, the course of the amplification reaction is determined

as a function of the cycle number and subsequently the n^{th} derivative is calculated. In this case exceeding certain extremes can be defined as exceeding a certain signal threshold value (EP Application No. 0016523.4). Hence this method of determining the threshold value is independent of the absolute signal strength of for example a fluorescence signal. Thus it is particularly suitable for those embodiments in which the target nucleic acid and reference nucleic acid are amplified in the same reaction vessel and are detected with the aid of different fluorescent labels. Methods have proven to be particularly suitable for the efficiency-corrected quantification of PCR products in which the maximum of the second derivative is determined as a measure for the signal threshold value.

[0059] The hybridization probes used for the methods according to the invention are usually single-stranded nucleic acids such as single-stranded DNA or RNA or derivatives thereof or alternatively PNAs which hybridize at the annealing temperature of the amplification reaction to the target nucleic acid. These oligonucleotides usually have a length of 20 to 100 nucleotides.

[0060] Depending on the detection format the label can be introduced on any ribose or phosphate group of the oligonucleotide. Labels at the 3' and 5' end of the nucleic acid molecule are preferred.

[0061] The type of label must be detectable in the real-time mode of the amplification reaction. This is for example in principle also (but not only) possible with the aid of labels that can be detected by NMR.

[0062] Methods are particularly preferred in which the amplified nucleic acids are detected with the aid of at least one fluorescent-labelled hybridization probe.

[0063] Many test procedures are possible for this. The following three detection formats have proven to be particularly suitable in connection with the present invention:

a) FRET hybridization probes

For this test format 2 single-stranded hybridization probes are used simultaneously which are complementary to adjacent sites of the same strand of the amplified target nucleic acid. Both probes are labelled with different fluorescent components. When excited with light of a suitable wavelength, a first component transfers the absorbed energy to the second component according to the principle of fluorescence resonance energy transfer such that a fluorescence emission of the second component can be measured when both hybridization probes bind to adjacent positions of the target molecule to be detected.

Alternatively it is possible to use a fluorescent-labelled primer and only one labelled oligonucleotide probe (Bernard et al., Analytical Biochemistry 235, p. 1001-107 (1998)).

b) TaqMan hybridization probes

A single-stranded hybridization probe is labelled with two components. When the first component is excited with light of a suitable wavelength, the absorbed energy is transferred to the second component, the so-called quencher, according to the principle of fluorescence resonance energy transfer. During the annealing step of the PCR reaction, the hybridization probe binds to the target DNA and is degraded by the 5'-3' exonuclease activity of the Taq polymerase during the subsequent elongation phase. As a result the excited fluorescent component and the quencher are spatially separated from one another and thus a fluorescence emission of the first component can be measured.

c) Molecular Beacons

These hybridization probes are also labelled with a first component and with a quencher, the labels preferably being located at both ends of the probe. As a result of the secondary structure of the probe, both components are in spatial proximity in solution. After hybridization to the target nucleic acid both components are separated from one another such that after excitation with light of a suitable wavelength the fluorescence emission of the first component can be measured (Lizardi et al., US 5,118,801).

[0064] In the described embodiments in which only the target nucleic acid or only the reference nucleic acid or an external standard is amplified in one reaction vessel in each case, the respective amplification product can also be detected according to the invention by a DNA binding dye which emits a corresponding fluorescence signal upon interaction with the double-stranded nucleic acid after excitation with light of a suitable wavelength. The dyes SybrGreen and SybrGold (Molecular Probes) have proven to be particularly suitable for this application. Intercalating dyes can alternatively be used.

[0065] A subject matter of the invention are also kits that contain appropriate agents to carry out the method according to the invention. According to the invention these agents are present in the kit in various compositions. A kit preferably contains reagents such as for example a reverse transcriptase for preparing a cDNA, DNA polymerase for the amplification reaction, specific primers for the amplification reaction and optionally also specific hybridization probes to detect the amplification product. As an alternative polymerases for a single-step RT-PCR reaction can be present in the kit. It is also possible that a kit according to the invention contains package inserts or disks containing files with previously

determined amplification efficiencies for defined amplification conditions. Finally the invention also concerns a kit which additionally contains further reagents for the synthesis and labelling of oligonucleotides such as fluorescent NHS-esters or fluorescent-labelled CPGs. Moreover a kit according to the invention can optionally also contain a DNA which can be used as an internal or external standard.

[0066] The invention is further elucidated by the following examples:

Example 1: Amplification of cytokeratin 20 (CK20) and porphobilinogen (PBGD) cDNAs

[0067] RNA was isolated from the cell line HT-29 (ATCC) using a HighPure-RNA Restriction Kit (Roche Diagnostics GmbH). After semi-quantitative spectrophotometric determination, the RNA concentration was adjusted to 100 ng/μl in RNA-free water. Three serial single dilutions were prepared from this with RNA concentrations of 10 ng, 1 ng and 100 pg/μl.

[0068] Total cDNA was prepared from these dilutions by reverse transcription under the following conditions:

1 x	AMV reverse transcription buffer
1 mM	of each deoxynucleoside triphosphate
0.0625 mM	randomized hexamers
10 u	AMV reverse transcriptase
10 μl	RNA
Ad. 20 μl	water

[0069] All mixtures were incubated for 10 minutes at 25°C, 30 minutes at 42°C and 5 minutes at 95°C for the cDNA synthesis. Subsequently they were cooled to 4°C. A sample containing 10 ng/μl HT29 RNA was used as a calibrator.

[0070] Afterwards the amplification reaction was carried out which was measured in real-time in the FRET HybProbe format on a LightCycler instrument (Roche Diagnostics GmbH). Each reaction mixture was amplified under the following conditions:

1 x	fast start DNA hybridization probes buffer
1 x	detection mix
2 μl	cDNA
Ad. 20 μl	water

[0071] The 1 x detection mix was composed of 0.5 μM forward and 0.5 μM reverse primers, each 0.2 μM fluorescein and LC-Red 640 labelled hybridization probes, 4 mM magnesium chloride and 0.005 % Brij-35.

[0072] Primers having SEQ ID NO: 1 and SEQ ID NO:2 were used to amplify a CK20 sequence. The CK20 product was detected using a fluorescein probe having SEQ ID NO:3 and a LC-Red 640 hybridization probe having SEQ ID NO:4. Primers having SEQ ID NO:5 and 6 were used to detect the PBGD sequence. PBGD was detected using a fluorescein-labelled hybridization probe having SEQ ID NO:7 and an LC-Red 640-labelled hybridization probe having SEQ ID NO:8.

[0073] The reaction mixtures were firstly incubated for 10 minutes at 95°C in the presence of 5 mM magnesium chloride for the amplification. The actual amplification reaction was carried out for 50 cycles according to the following scheme:

10 sec. 95°C
10 sec. 60°C
5 sec. 72°C

[0074] After each incubation at 60°C a fluorescence measurement was carried out according to the manufacturer's instructions. The signal threshold value (Cp value) was determined as the maximum of the 2nd derivative of the amplification reaction as a function of the cycle number.

Example 2: Determination of the efficiency of the amplification of CK20 and PBGD

[0075] A function was established to determine the efficiency in which the cycle number Cp determined for the respective concentration was determined as a function of the decadic logarithm of the RNA concentration used.

[0076] A linear function was calculated from this function by regression analysis with the aid of the LightCycler soft-

ware. Starting from this function the efficiency was determined according to the equation

$$\text{efficiency} = 10^{-1/a}$$

wherein a is the gradient (1st derivative) of the determined regression line.

Table 2:

Conc (ng)	Log (ng)	Cp-CK20	Cp-PBGD
0.1	-1.0	35.73	38.73
1	0.0	30.13	33.59
10	1.0	24.20	28.63
Efficiency:		1.491	1.578
Cp: measured cycle number			

[0077] The results obtained for CK20 and PBGD are shown in table 2. The result shows that on the one hand the efficiencies are considerably different from 2.00 i.e. a doubling of the target nucleic acid does not take place with each PCR cycle. On the other hand, the result shows that the efficiencies of the amplification of CK20 and PBGD differ significantly from one another under otherwise identical conditions.

Example 3: Determination of the original ratio of target nucleic acid and reference nucleic acid with and without correction of the amplification efficiency

[0078] Under the conditions described in example 1 the ratio determined of the original amount of CK20 and PBGD should be independent of the respective amplified concentration of the sample material used. Hence the determination of the ratio for various amounts of sample RNA was used to check the effect of an efficiency correction on the basis of the measured values that were obtained. In this case the ratio of CK20 to PBGD was determined according to the invention according to equation (5). On the one hand, the ratio was determined using the efficiencies obtained from example 2 and on the other hand with an assumed amplification efficiency of 2.00 for CK20 and for PGD. The results are shown in table 3:

Table 3:

HT29 (ng)	CP CK20	Cp PBGD	$N(T)_0/N(R)_0$ Efficiency=2	$N(T)_0/N(R)_0$ Efficiency corrected
0.1ng	35.73	38.73	8.00	29.68
1ng	30.13	33.59	11.00	26.66
10ng	24.20	28.63	21.56	29.66
M:			13.52	28.66
SD:			7.12	1.74
%CV:			52.7%	6.1%
Cp= measured cycle number M= mean SD= standard deviation %CV= coefficient of variation				

[0079] As can be seen from the table, the efficiency-corrected values calculated for the ratio of $N(T)_0/N(R)_0$ have a significantly lower standard deviation for the various amounts of sample RNA than the uncorrected values and a coefficient of variation of 6.1 % compared to 52.7 %.

Example 4: Efficiency-correction when using a calibrator

[0080] Analogously to examples 1 and 2 amplification reactions were carried out in the presence of 10 mM magnesium chloride. In this case an efficiency of 1.491 was determined for CK20 and an efficiency of 1.578 was determined for PBGD. In addition the Cp values of a calibrator sample containing an unknown amount of HT-29 RNA was determined at 5 mM and 10 mM magnesium chloride. The measured data were used to determine the quotients of the ratios of CK20 to PBGD between the samples analysed in each case and the appropriate calibrator according to equation (7). This determination was carried out on the one hand with an assumed efficiency of 2 for the amplification of CK20 and PBGD as well as, on the other hand, with the aid of experimentally determined amplification efficiencies. The result is shown in table 4.

Table 4:

MgCl ₂	HT29 (ng)	Cp CK20	Cp PBGD	T:R/C Efficiency	T:R/C Efficiency corrected
5mM	0.1ng	36.59	39.09	0.76	0.92
5mM	1ng	30.60	32.60	0.54	0.72
5mM	10ng	25.19	27.95	0.91	0.95
calibrator	Cal.	24.78	27.67	1.00	1.00
10mM	0.1ng	35.73	38.73	0.39	1.04
10mM	1ng	30.13	33.59	0.53	0.93
10mM	10ng	24.20	28.63	1.04	1.04
calibrator	Cal.	24.01	28.38	1.00	1.00
M:				0.70	0.93
SD:				0.26	0.10
%CV:				36.7%	11.1%

T:R/C =

$$\begin{aligned}
 \text{Cp} &= \text{measured cycle number} \\
 M &= m \frac{N(T)_{0A}}{N(T)_{0K}} \\
 \text{SD} &= st \frac{N(R)_{0A}}{N(R)_{0K}} \\
 \% CV &= cc \frac{N(R)_{0A}}{N(R)_{0K}}
 \end{aligned}$$

[0081] As can be seen in table 4, the efficiency-corrected values have a lower standard deviation (0.10) as well as a three-fold lower coefficient of variation than the T:R/C values with an assumed PCR efficiency of 2.00. This result shows that an efficiency correction according to the invention is also advantageous in quantifications in which a standardization with the aid of calibrators has already been carried out.

Example 5: Absolute quantification of plasmid DNA

[0082] A decadic dilution series of a plasmid containing the PSA gene of 10⁹ to 10² copies was prepared for this purpose. At the same time a second decadic dilution series with a plasmid containing the gene for TNF (tumour necrosis factor) with an unknown copy number of plasmid DNA was prepared. Afterwards the PSA reaction mixtures were amplified on a LightCycler (Roche Diagnostics) under standard conditions using the primers having SEQ ID NO:9 and 10 and the TNF reaction mixtures were amplified using the primers having SEQ ID NO:11 and 12 (Roche Diagnostics LightCycler SybrGreen Mastermix, 5 mM final concentration MgCl₂, 0.5 μM final concentration of each primer). The amplification was measured in real-time using the DNA binding agent SybrGreenI (Molecular Probes) under standard conditions in which the evaluation was carried out according to the manufacturer's instructions in the second derivative mode.

[0083] The original copy number of the TNF plasmid was determined in two different ways on the basis of the obtained

data.

[0084] On the one hand a calibration line based on the PSA amplification was generated assuming the same amplification efficiency for PSA and TNF.

[0085] On the other hand the original copy number was determined according to formula (8). Analogously to example 2 the amplification efficiency for PSA and TNF was determined by calculating a regression line according to the formula

$$E = 10^{-1/a}$$

wherein a denotes the increase (1st derivative) of the calculated regression line. In this case an amplification efficiency of 2.03 was determined for PSA and an amplification efficiency of 2.13 was determined for TNF.

[0086] The results of the two different quantification procedures are shown in table 5. A so-called dilution check was carried out as a measure for the accuracy of the determination. The values denoted dilution check are calculated from the quotients of the copy numbers measured for the respective dilution of two dilution mixtures that differ from one another by a factor of 10. Thus a value of 10.00 would be expected as the ideal value.

Table 5

Dilution	Not efficiency corrected		Efficiency corrected	
	Determined copy number per dilution	Dilution check	Determined copy number per dilution	Dilution check
1	30826128	10.10	27728632	12.12
10 ⁻¹	3053000	13.82	2287050	14.98
10 ⁻²	220900	7.19	152643	7.94
10 ⁻³	30710	11.61	19227	13.89
10 ⁻⁴	2646	8.55	1384	9.52
10 ⁻⁵	309.5	7.61	145.4	8.76
10 ⁻⁶	40.66	3.86	16.6	3.84
10 ⁻⁷	10.54		4.3	
Mean:		8.96		10.16

[0087] As the result of the dilution check from table 5 shows, the mean of the efficiency-corrected data results in a value of 10.16, whereas the mean of non-efficiency-corrected data results in a value of 8.96 which is considerably further away from the ideal value of 10.00. From this it follows that an efficiency correction is also advantageous for embodiments in which an absolute quantification of nucleic acids with the aid of PCR is carried out.

SEQUENCE PROTOCOL

5 <110> Roche Diagnostics GmbH

<120> Method for the efficiency-corrected real-time quantification of
nucleic acids

10 <130> 539400DE

<140>

<141>

<160> 12

15 <170> PatentIn Ver. 2.1

<210> 1

<211> 20

<212> DNA

<213> Homo sapiens

20 <400> 1

atcaagcagt ggtacgaaac 20

25 <210> 2

<211> 18

<212> DNA

<213> Homo sapiens

<400> 2

30 aggacacacc gagcattt 18

<210> 3

<211> 25

<212> DNA

<213> Homo sapiens

35 <400> 3

attacagaca aattgaagag ctgcg 25

40 <210> 4

<211> 25

<212> DNA

<213> Homo sapiens

<400> 4

45 agtcagatta aggatgctca actgc 25

<210> 5

<211> 19

<212> DNA

<213> Homo sapiens

50 <400> 5

gcggagccat gtctggtaa 19

55

5	<p><210> 6 <211> 20 <212> DNA <213> Homo sapiens</p> <p><400> 6 ccagggtacg aggctttcaa</p>	20
10	<p><210> 7 <211> 24 <212> DNA <213> Homo sapiens</p> <p><400> 7 gagagtgatt cgcgtaggta cccg</p>	24
20	<p><210> 8 <211> 25 <212> DNA <213> Homo sapiens</p> <p><400> 8 agagccagct tgctcgcata cagac</p>	25
25	<p><210> 9 <211> 22 <212> DNA <213> Homo sapiens</p> <p><400> 9 gaggagttct tgacccccaaa ga</p>	22
35	<p><210> 10 <211> 18 <212> DNA <213> Homo sapiens</p> <p><400> 10 tccagcgctc agcacaca</p>	18
40	<p><210> 11 <211> 20 <212> DNA <213> Homo sapiens</p> <p><400> 11 cctgccccaa tccctttatt</p>	20
50	<p><210> 12 <211> 21 <212> DNA <213> Homo sapiens</p> <p><400> 12 ggtttcgaag tggtaggtctt g</p>	21
55		

Claims

1. Method for the quantification of a target nucleic acid in a sample comprising the following steps:

- 5 a) determination of the amplification efficiency of the target nucleic acid under defined amplification conditions
 b) amplification of the target nucleic acid contained in the sample under the same defined reaction conditions
 c) measuring the amplification in real-time
 d) quantification of the original amount of target nucleic acid in the sample by correction of the original amount
 10 derived from step c) with the aid of the determined amplification efficiency.

2. Method as claimed in claim 1, wherein the efficiency of the amplification is determined by

- 15 a) preparing a dilution series of the target nucleic acid
 b) amplifying the target nucleic acid under defined reaction conditions as claimed in claim 1, the amplification of the nucleic acid being measured in real time
 c) determining a defined threshold value
 d) determining the cycle number at which the signal threshold value is exceeded for each dilution
 e) determining a logarithmic linear function of the copy number of target nucleic acid used for the amplification
 20 as a function of the cycle number at which the signal threshold value is exceeded and
 f) calculating the amplification efficiency E according to

$$E = G^{-a}$$

25 wherein a is determined as the first derivative of the function determined in step e) and G is the base number of the logarithm.

3. Method as claimed in claim 1, wherein the efficiency of the amplification is determined by

- 30 a) preparing a dilution series of the target nucleic acid
 b) amplifying the target nucleic acid under defined reaction conditions as claimed in claim 1, the amplification of the nucleic acid being measured in real time
 c) determining a defined signal threshold value
 35 d) determining the cycle number at which the signal threshold value is exceeded for each dilution
 e) determining a linear function of the cycle number determined in step d) as a function of a logarithm of the copy number of target nucleic acid used for the amplification and
 f) calculating the amplification efficiency E according to

$$E = G^{-1/a}$$

40 wherein a is determined as the first derivative of the function determined in step e) and G is the base number of the logarithm.

45 4. Method as claimed in claim 1 wherein the efficiency of the amplification is determined by

- 50 a) preparing a dilution series of the target nucleic acid
 b) amplifying the target nucleic acid under defined reaction conditions as claimed in claim 1, the amplification of the nucleic acid being measured in real time
 c) determining a defined signal threshold value
 d) determining the cycle number at which the signal threshold value is exceeded for each dilution
 e) determining the amplification efficiency as a function of the amount of target nucleic acid.

55 5. Method for the quantification of a target nucleic acid in a sample relative to a reference nucleic acid comprising the following steps:

- a) determination of the amplification efficiencies of the target nucleic acid and of the reference nucleic acid

under defined amplification conditions

b) amplification of the target nucleic acid contained in the sample and of the reference nucleic acid contained in the sample under the same defined reaction conditions

c) measuring the amplification of the target nucleic acid and of the reference nucleic acid in real time

d) calculation of the original ratio of target nucleic acid and reference nucleic acid in the sample by correction of the ratio derived from step c) with the aid of the amplification efficiencies determined in step a).

6. Method as claimed in claim 5, wherein steps b) - d) are additionally carried out using a calibrator sample and subsequently the ratio of the quotients determined for the sample and for the calibrator sample is determined as a measure for the original amount of target nucleic acid in the sample.

7. Method for the quantification of a target nucleic acid in a sample relative to a reference nucleic acid comprising the following steps:

a) determination of the amplification efficiencies of the target nucleic acid and of the reference nucleic acid under defined amplification conditions

b) amplification of the target nucleic acid contained in the sample as well as of the reference nucleic acid contained in the sample under the same defined amplification conditions.

c) measurement of the amplification of the target nucleic acid and of the reference nucleic acid in real time.

d) determining a defined signal threshold value

e) determining the cycle numbers at which the signal threshold value is exceeded in each case during the amplification of target nucleic acid and reference nucleic acid

f) calculating the original ratio of target nucleic acid and of reference nucleic acid in the sample according to the following formula

$$N(T)_0 / N(R)_0 = E(R)^{n(R)} / E(T)^{n(T)},$$

in which

$N(T)_0$ = the original amount of target DNA present in the sample

$N(R)_0$ = the original amount of reference DNA present in the sample

$E(R)$ = the amplification efficiency of the reference nucleic acid

$n(R)$ = the cycle number of the reference nucleic acid measured in step e)

$E(T)$ = the amplification efficiency of the target nucleic acid

$n(T)$ = the cycle number of the target nucleic acid measured in step e).

8. Method as claimed in claim 6, wherein steps b), c), e) and f) are additionally carried out using a calibrator sample and subsequently the ratio of the quotients determined for the sample and for the calibrator sample is determined as a measure for the original amount of target nucleic acid in the sample.

9. Method as claimed in claims 5-8, wherein the determination of the amplification efficiency of the target nucleic acid and the determination of the amplification efficiency of the reference nucleic acid is carried out by a method as claimed in claims 2-4.

10. Method as claimed in claims 5-9, wherein the real-time measurement of the amplification of target nucleic acid and reference nucleic acid in a sample is carried out in separate reaction vessels.

11. Method as claimed in claims 5-9, wherein the real-time measurement of the amplification of target nucleic acid and reference nucleic acid in a sample is carried out in the same reaction vessel using differently labelled hybridization probes.

12. Method for the quantification of a target nucleic acid in a sample comprising the following steps:

a) determination of the amplification efficiencies of the target nucleic acid and of an internal or external standard under defined amplification conditions

- b) amplification of the target nucleic acid contained in the sample and of the internal or external standard under the same defined reaction conditions
- c) measuring the amplification of the target nucleic acid and of the standard in real time
- d) calculating the original copy number in the sample by correction of the copy number derived from step c) with the aid of the amplification efficiencies determined in step a).

13. Method for the quantification of a target nucleic acid in a sample comprising the following steps:

- a) determination of the amplification efficiencies of the target nucleic acid and of an internal or external standard under defined amplification conditions
- b) amplification of the target nucleic acid contained in the sample as well as of the internal or external standard under the same defined reaction conditions.
- c) measurement of the amplification of the target nucleic acid and of the standard in real time.
- d) determining a defined signal threshold value
- e) determining the cycle numbers at which the signal threshold value is exceeded during the amplification of target nucleic acid and standard
- f) determining the original copy number $N(T)_0$ of target nucleic acid in the sample according to the formula

$$N(T)_0 = N(S)_0 \cdot E(S)^{n(S)} / E(T)^{n(T)},$$

in which

- $N(S)_0$ = the original amount of standard used
- $E(S)$ = the amplification efficiency of the standard
- $n(S)$ = the cycle number of the standard measured in step e)
- $E(T)$ = the amplification efficiency of the target nucleic acid
- $n(T)$ = the cycle number of the target nucleic acid measured in step e).

14. Method as claimed in claims 12-13, wherein the amplification efficiencies are determined as claimed in claims 2-4.

15. Method as claimed in claims 12-14 using an internal standard, wherein real-time measurement of the amplification of the target nucleic acid and internal standard is carried out with differently labelled hybridization probes.

16. Method as claimed in claims 1-15, wherein the amplified nucleic acids are detected with the aid of at least one fluorescent-labelled hybridization probe.

17. Method as claimed in claim 16, wherein the amplified nucleic acids are detected with the aid of FRET hybridization probes, molecular beacons or TaqMan probes.

18. Method as claimed in claims 1-10 or 12-14, wherein the amplified nucleic acids are detected with the aid of a DNA-binding dye, preferably with SybrGreen I.

19. Kit containing agents to carry out the method as claimed in claims 1-18.